



Certificate of Conformity No 213-02573

<i>Object</i>	Power quality instrument SONEL PQM-750 Product classification: PQI-A-FI1
<i>Applicant</i>	SONEL S.A. ul. Stanisława Wokulskiego 11 58-100 Świdnica Poland
<i>Requirements</i>	Certification IEC 61000-4-30:2015+AMD1:2021, Class A Testing according to IEC 62586-2:2017+AMD1:2021 (functional tests) and IEC 61180:2016 + IEC 61000-4-4:2012 (EMC immunity tests)
<i>Confirmation</i>	See page 2
<i>Date of Examination</i>	19 February to 21 May 2024
3003 Bern-Wabern, 5 June 2024	
<i>For the Examination</i>	Christian Santschi
<i>Approved by</i>	Dr Cédric Blaser, Head of Laboratory Laboratory Electrical Energy and Power

Certificate of Conformity No 213-02573

Type description

Type SONEL PQM-750
Classification PQI-A-FI1 as per IEC 62586-1:2017
Class A
Fixed Installed
Indoor application
EMC environment G
Number of phases 3
Connection to power lines..... 3P+N (4 wires)
Declared input voltage (U_{din}) 230 V (50 Hz), 120 V (60 Hz)
Nominal value of current input (I_n)..... 5 A
Fundamental frequency (f_n)..... 50 Hz, 60 Hz
Hardware version..... HWa
Firmware version 0.19 C3
Evaluation software SONEL ANALYSIS, Version 4.6.9, Build 12

Conformity assessment against IEC 62586-2:2017+AMD1:2021

Power quality parameter	Sub-clause	Compliance 230 V / 5 A ; 50 Hz	Compliance 120 V / 5 A ; 60 Hz
Power frequency	6.1	Yes	Yes
Magnitude of supply voltage	6.2	Yes	Yes
Flicker, class F1	6.3	Yes	Yes
Supply voltage interruptions, dips and swells	6.4	Yes	Yes
Supply voltage unbalance	6.5	Yes	Yes
Voltage harmonics	6.6	Yes	Yes
Voltage interharmonics	6.7	Yes	Yes
Mains signalling voltages on the supply voltage	6.8	Yes	Yes
Measurement of under-deviation and over-deviation parameters	6.9	N.A.	N.A.
Flagging	6.10	Yes	Yes
Clock uncertainty testing	6.11	Yes	Yes
Variations due to external influence quantities	6.12	Yes	Yes
Rapid voltage changes (RVC)	6.13	Yes	Yes
Magnitude of current	6.14	Yes	Yes
Harmonic current	6.15	Yes	Yes
Interharmonic currents	6.16	Yes	Yes
Current unbalance	6.17	Yes	Yes
Calculation of measurement uncertainty and operating uncertainty	8	Yes	Yes

Certificate of Conformity No 213-02573

EMC immunity test according to IEC 61000-4-30:2015+AMD1:2021

Test requirement	Normative reference	Compliance 230 V / 5 A ; 50 Hz	Compliance 120 V / 5 A ; 60 Hz
Transient voltages	IEC 61180:2016	Yes	Yes
Electrical fast transient (burst)	IEC 61000-4-4:2012	Yes	Yes

Test reports

Topic	Issuing Laboratory	Report number	Report date
Power quality parameter	METAS	213-02568	2024-05-31
EMC immunity	METAS	218-03525	2024-06-04

Full name and addresses of issuing laboratories

- Federal Institute of Metrology METAS, Lindenweg 50, 3003 Bern-Wabern, Switzerland

Certificate of Conformity No 213-02573

Note

The Federal Institute of Metrology (METAS) is Switzerland's national metrology institute.

Its tasks and activities arise from the Federal Law on the Institute of Metrology and the Federal Law on Metrology and its implementation provisions. METAS's principal tasks include making available the internationally recognised units of measurement in compliance with the International System of Units (SI), comparing them with other national metrology institutes and disseminating them.

METAS thus plays the leading role in the Swiss metrological infrastructure and ensures the traceability of measurement results in legal metrology, at accredited calibration and test laboratories and laboratories in industry, research and administration through the accuracy and reliability of its services.

METAS operates a comprehensive quality management system covering the entire scope of activities in the metrology area. The METAS management system satisfies the statutory requirements and the general requirements for the competence of testing and calibration laboratories defined in standard ISO/IEC 17025:2017 in their entirety. Therefore, results from METAS provide traceability as required by ISO/IEC 17025:2017 subclause A.3.1.